

### Instrument Field Response Check Log

**1. Instrument Information<sup>1</sup>**

Ratemeter: Make/Model: Ludlum 2241-2 Serial No. 206098 Cal. Due Date: 09/01/16  
 Detector 1: Make/Model: Ludlum 44-10 Serial No. PR112642  
 Bicron MicroRem Meter: Serial No. \_\_\_\_\_ Cal. Due Date: \_\_\_\_\_

**2. Check Source Information:**

Source 1 Isotope: Th-232 Serial No.: 111 Activity: 0.1 units: NC Assay Date: 12/30/10  
 Response Acceptance Range (+/-20%): uRem/hr +20% \_\_\_\_\_ uRem/hr -20% \_\_\_\_\_ net cpm + 20% 5348 net cpm -20% 35866  
 Source 2 Isotope: Cs-137 Serial No.: 119E23-12 Activity: 0.02 units: NC Assay Date: NA  
 Response Acceptance Range (+/-20%): uRem/hr +20% \_\_\_\_\_ uRem/hr -20% \_\_\_\_\_ net cpm + 20% 13273 net cpm -20% 8899

**3. Technician/Worker Performing Checks:**

Name: J. Edwards Title: RCT Date: 12/17/15 Time: 0815

**4. Site or Location:**

Site/Job: Area 5.5-5.6 Location Description: woods  
 GPS Coordinates (when required): X-Coord: \_\_\_\_\_ Y-Coord: \_\_\_\_\_

Instrument Field Response <sup>2</sup>					Use Acceptance Criteria				Remarks	
Meter	Bkg Cnt Time	Bkg Counts (cpm) or uRem/hr	Source Cnt Time	Source Response (gross cpm or uRem/hr)	+/- 20% source gross cpm or uRem/hr (Y/N)	Inst. Calib. current (Y/N)	Battery Check (Y/N)	Time Of check	Ambient Temp. (°F)	Initials and Comments (add'l info: inst. Condition, etc.)
Ratemeter	1min	9274 cpm	1min	44808 cpm	Y	Y	Y	0819	49.4°	Th-232 DE
Ratemeter	"	"	1min	11109 cpm	Y	Y	Y	0823	49.4°	Cs-137 DE
Ratemeter	1min	9433 cpm	1min	46391 cpm	Y	Y	Y	1033	50.1°	Th-232 DE
Ratemeter	"	"	1min	11440 cpm	Y	Y	Y	1040	50.1°	Cs-137 DE
Ratemeter	1min	7004 cpm	1min	44433 cpm	Y	Y	Y	1400	43.5°	Th-232 TB
Ratemeter	"	"	1min	9743 cpm	Y	Y	Y	1400	43.5°	Cs-137 TB
Bicron	NA	5 pps/hr	NA	30 pps/hr	Y	Y	Y	1035	50.3°	Th-232 DE
Bicron	NA	N/A	NA	N/A	Y	Y	Y			N/A
Bicron	NA	5 uRem/hr	NA	30 uRem/hr	Y	Y	Y	1400	43.5°	Th-232 TB

1. Instrument designated check source is listed on calibration sticker. Record check source response (net cpm) prior to field deployment for all check sources being used.  
 2. Source and Background count rate should be determined from the average of three static counts at the same location. Repeat counts should be within 20%. If count rate diverges significantly, perform additional counts to evaluate instrument stability