



# Certificate of Calibration

2015-5999-1

243 Root Street Suite 100  
Olean, NY 14760

Customer		Instrument	
Customer Name: MJW Corporation		Manufacturer: Bicon	
Address: 15 Hazelwood Drive Suite 112 Amherst, NY 14228		Model: Micro Rem	SN: B693G
Customer PO#		Detector Manufacturer: Bicon	
Work Order: 2015-5999		Det. Model: Internal Scintillator	SN: n/a
Calibration Method: Electronic(*) and Source		Procedure: MCP-01	
Instrument Received: <input checked="" type="checkbox"/> Within Tolerance <input type="checkbox"/> Out of Tolerance <input type="checkbox"/> Repairs Required <input type="checkbox"/> Other (See Comments)			
<input checked="" type="checkbox"/> Geotropism <input checked="" type="checkbox"/> Meter Zero <input checked="" type="checkbox"/> Mech. Ck <input checked="" type="checkbox"/> HV Readout <input checked="" type="checkbox"/> Battery Check <input type="checkbox"/> Reset			
<input type="checkbox"/> Audio <input type="checkbox"/> Window Status <input type="checkbox"/> FS Response <input checked="" type="checkbox"/> Linearity <input type="checkbox"/> Alarm Set			
Temperature: 70 F		Humidity: 50%	Pressure: 28.7 in Hg
Altitude: 1450 ft			

Instrument Calibration									
Multiplier/Range	Calibration Point	Instrument Response			Multiplier/Range	Calibration Point	Instrument Response		
		As Found	As Left	Tolerance			As Found	As Left	Tolerance
X 0.1*	4 uR/hr	3.9 uR/hr	3.9 uR/hr	3.4-4.6	X 0.1*	16 uR/hr	16 uR/hr	16 uR/hr	14.4-17.6
X 1*	40 uR/hr	40 uR/hr	40 uR/hr	34-46	X 1	160 uR/hr	160 uR/hr	160 uR/hr	144-176
X 10	400 uR/hr	400 uR/hr	400 uR/hr	340-460	X 10	1600 uR/hr	1600 uR/hr	1600 uR/hr	1440-1760
X 100	4000 uR/hr	4000 uR/hr	4000 uR/hr	3400-4600	X 100	16000 uR/hr	16000 uR/hr	16000 uR/hr	14400-17600
X 1K	40000 uR/hr	40000 uR/hr	40000 uR/hr	34000-46000	X 1K	160000 uR/hr	162000 uR/hr	162000 uR/hr	144000-176000
X 0.1*	400 cpm =	3.9 uR/hr	3.9 uR/hr	For Reference	X 0.1*	1600 cpm =	16 uR/hr	16 uR/hr	For Reference
X 1*	4 kcpm =	40 uR/hr	40 uR/hr	For Reference	X 1*	16 kcpm =	160 uR/hr	160 uR/hr	For Reference

### Electronic Instruments

Pulser: 500-2 sn# 220106

### Sources

Isotope	Serial#	Type	Activity	Response	Efficiency	CF	Distance
Cs137	7773CM	Cell I Gamma Irradiator	1.581 Ci				

### Instrument Status

Disc/Sens: 610 mV

### Statement of Certification

MJW Technical Services, Inc certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology or to the calibrator facilities of other International Standards organization members, or have been derived from accepted values of natural physical constants or have been derived by the ration type of calibration techniques. The calibration system conforms to the requirements ISO/IEC 17025 and ANSI N323. The Instrument listed above was inspected prior to shipment and it met all the manufacturer's published operating specifications. (MJW Technical Services is not responsible for damage incurred during shipment or use of this instrument).

Calibration Technician: (AL)	QA Technician:	Date: 5-6-15
Calibration Date: 05/05/2015	Calibration Due: 05/05/2016	



# Certificate of Calibration

2015-6089-2

243 Root Street Suite 100  
Olean, NY 14760

Customer		Instrument	
Customer Name: MJW Technical Services		Manufacturer: Bicron	
Address: 243 Root Street Suite 100 Olean, NY 14760		Model: Micro Rem	SN: 1487
Customer PO#		Detector Manufacturer: Bicron	
Work Order: 2015-6089		Det. Model: Internal Scintillator	SN: n/a
Calibration Method: Electronic(*) and Source		Procedure: MCP-01	
Instrument Received: <input type="checkbox"/> Within Tolerance <input checked="" type="checkbox"/> Out of Tolerance <input type="checkbox"/> Repairs Required <input type="checkbox"/> Other (See Comments)			
<input checked="" type="checkbox"/> Geotropism <input checked="" type="checkbox"/> Meter Zero <input checked="" type="checkbox"/> Mech. Ck <input checked="" type="checkbox"/> HV Readout <input checked="" type="checkbox"/> Battery Check <input checked="" type="checkbox"/> Reset			
<input type="checkbox"/> Audio <input type="checkbox"/> Window Status <input type="checkbox"/> FS Response <input checked="" type="checkbox"/> Linearity <input type="checkbox"/> Alarm Set			
Temperature: 71.2 F		Humidity: 57%	
		Pressure: 28.5 in Hg	
		Altitude: 1450 ft	

Instrument Calibration									
Multiplier/Range	Calibration Point	Instrument Response			Multiplier/Range	Calibration Point	Instrument Response		
		As Found	As Left	Tolerance			As Found	As Left	Tolerance
X 0.1*	4 uR/hr	3.8 uR/hr	3.8 uR/hr	3.4-4.6	X 0.1*	16 uR/hr	16.2 uR/hr	16.2 uR/hr	14.4-17.6
X 1*	40 uR/hr	40 uR/hr	40 uR/hr	34-46	X 1	160 uR/hr	<b>130 uR/hr</b>	160 uR/hr	144-176
X 10	400 uR/hr	350 uR/hr	400 uR/hr	340-460	X 10	1600 uR/hr	<b>1420 uR/hr</b>	1600 uR/hr	1440-1760
X 100	4000 uR/hr	3500 uR/hr	4000 uR/hr	3400-4600	X 100	16000 uR/hr	14500 uR/hr	16000 uR/hr	14400-17600
X 1K	40000 uR/hr	34000 uR/hr	41000 uR/hr	34000-46000	X 1K	160000 uR/hr	<b>142000 uR/hr</b>	160000 uR/hr	144000-176000
X 0.1*	400 cpm =	400 uR/hr	400 uR/hr	For Reference	X 0.1*	1600 cpm =	1600 uR/hr	1600 uR/hr	For Reference
X 1*	4 kcpm =	4000 uR/hr	4000 uR/hr	For Reference	X 1*	16 kcpm =	160000 uR/hr	160000 uR/hr	For Reference

### Electronic Instruments

Pulser: 500-2 sn# 220099

### Sources

Isotope	Serial#	Type	Activity	Response	Efficiency	CF	Distance
Cs137	7773CM	Cell 1 Gamma Irradiator	1.577 Ci				

### Instrument Status

Disc/Sens: 650 mV

### Statement of Certification

MJW Technical Services, Inc certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology or to the calibrator facilities of other International Standards organization members, or have been derived from accepted values of natural physical constants or have been derived by the ration type of calibration techniques. The calibration system conforms to the requirements ISO/IEC 17025 and ANSI N323. The Instrument listed above was inspected prior to shipment and it met all the manufacturer's published operating specifications. (MJW Technical Services is not responsible for damage incurred during shipment or use of this instrument).

Calibration Technician: (GH)		QA Technician:		Date:	6/19/15
Calibration Date:	06/18/2015	Calibration Due:	06/18/2016		



# Certificate of Calibration

2015-6178-4

243 Root Street Suite 100  
Olean, NY 14760

Customer		Instrument	
Customer Name: MJW Corporation		Manufacturer: Bicon	
Address: 15 Hazelwood Drive Suite 112 Amherst, NY 14228		Model: Micro Rem	SN: A224U
Customer PO#		Detector Manufacturer: Bicon	
Work Order: 2015-6178		Det. Model: Internal Scintillator	SN: n/a
Calibration Method: Electronic(*) and Source		Procedure: MCP-01	
Instrument Received: <input checked="" type="checkbox"/> Within Tolerance <input type="checkbox"/> Out of Tolerance <input type="checkbox"/> Repairs Required <input type="checkbox"/> Other (See Comments)			
<input checked="" type="checkbox"/> Geotropism <input checked="" type="checkbox"/> Meter Zero <input checked="" type="checkbox"/> Mech. Ck <input checked="" type="checkbox"/> HV Readout <input checked="" type="checkbox"/> Battery Check <input type="checkbox"/> Reset			
<input type="checkbox"/> Audio <input type="checkbox"/> Window Status <input type="checkbox"/> FS Response <input checked="" type="checkbox"/> Linearity <input type="checkbox"/> Alarm Set			
Temperature: 72 F		Humidity: 50%	Pressure: 28.4 in Hg
Altitude: 1450 ft			

## Instrument Calibration

Multiplier/Range	Calibration Point	Instrument Response			Multiplier/Range	Calibration Point	Instrument Response		
		As Found	As Left	Tolerance			As Found	As Left	Tolerance
X 0.1*	4 uR/hr	3.9 uR/hr	3.9 uR/hr	3.4-4.6	X 0.1*	16 uR/hr	16.1 uR/hr	16.1 uR/hr	14.4-17.6
X 1*	40 uR/hr	40 uR/hr	40 uR/hr	34-46	X 1	160 uR/hr	160 uR/hr	160 uR/hr	144-176
X 10	400 uR/hr	400 uR/hr	400 uR/hr	340-460	X 10	1600 uR/hr	1600 uR/hr	1600 uR/hr	1440-1760
X 100	4000 uR/hr	4000 uR/hr	4000 uR/hr	3400-4600	X 100	16000 uR/hr	16200 uR/hr	16200 uR/hr	14400-17600
X 1K	40000 uR/hr	40000 uR/hr	40000 uR/hr	34000-46000	X 1K	160000 uR/hr	160000 uR/hr	160000 uR/hr	144000-176000
X 0.1*	400 cpm =	3.9 uR/hr	3.9 uR/hr	For Reference	X 0.1*	1600 cpm =	16.1 uR/hr	16.1 uR/hr	For Reference
X 1*	4 kcpm =	40 uR/hr	40 uR/hr	For Reference	X 1*	16 kcpm =	160 uR/hr	160 uR/hr	For Reference

## Electronic Instruments

Pulser: 500-2 sn# 220106

## Sources

Isotope	Serial#	Type	Activity	Response	Efficiency	CF	Distance
Cs137	7773CM	Cell I Gamma Irradiator	1.572 Ci				

## Instrument Status

Disc/Sens: 620 mV

## Statement of Certification

MJW Technical Services, Inc certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology or to the calibrator facilities of other International Standards organization members, or have been derived from accepted values of natural physical constants or have been derived by the ratio type of calibration techniques. The calibration system conforms to the requirements ISO/IEC 17025 and ANSI N323. The instrument listed above was inspected prior to shipment and it met all the manufacturer's published operating specifications. (MJW Technical Services is not responsible for damage incurred during shipment or use of this instrument).

Calibration Technician: (AL)	QA Technician:	Date: 8/19/15
Calibration Date: 08/04/2015	Calibration Due: 08/04/2016	



# Certificate of Calibration

2015-6241-1

243 Root Street Suite 100  
Olean, NY 14760

Customer		Instrument	
Customer Name: MJW Corporation		Manufacturer: Ludlum Measurements	
Address: 15 Hazelwood Drive Suite 112 Amherst, NY 14228		Model: 2241-2                      SN: 206098	
Customer PO#                      Work Order: 2015-6241		Detector Manufacturer: Ludlum Measurements	
Calibration Method: Electronic(*)		Det. Model: 44-10                      SN: PR112642                      1-Red	
		Procedure: SOP-06	

Instrument Received:     Within Tolerance     Out of Tolerance     Repairs Required     Other (See Comments)

Geotropism     Meter Zero     Mech. Ck     HV Readout     Battery Check     Reset

Audio     Window Status     FS Response     Linearity     Alarm Set

Temperature: 71 F                      Humidity: 52%                      Pressure: 28.5 in Hg                      Altitude: 1450 ft

### Instrument Calibration

Multiplier/Range	Calibration Point	Instrument Response			Multiplier/Range	Calibration Point	Instrument Response		
		As Found	As Left	Tolerance			As Found	As Left	Tolerance
Ratemeter Mode*	40 cpm	40 cpm	40 cpm	36-44	Ratemeter Mode*	400 cpm	399 cpm	399 cpm	360-440
Ratemeter Mode*	4 kcpm	3.99 kcpm	3.99 kcpm	3.6-4.4	Ratemeter Mode*	40 kcpm	39.9 kcpm	39.9 kcpm	36-44
Ratemeter Mode*	400 kcpm	399 kcpm	399 kcpm	360-440	Scaler Mode*	40 cpm	40 cpm	40 cpm	36-44
Scaler Mode*	400 cpm	400 cpm	400 cpm	360-440	Scaler Mode*	4 kcpm	3.993 kcpm	3.993 kcpm	3.6-4.4
Scaler Mode*	40 kcpm	39.929 kcpm	39.929 kcpm	36-44	Scaler Mode*	400 kcpm	399.944 kcpm	399.944 kcpm	360-440

### Electronic Instruments

Pulser: 500-2 sn# 220106

### Sources

Isotope	Serial#	Type	Activity	Response	Efficiency	CF	Distance
Am241	UB992	Y	2344511.523 dpm	159080	6.8 %	14.71	0 inch
Cs137	UB994	Y	2381013.672 dpm	220200	9.2 %	10.87	0 inch

### Instrument Status

Disc/Sens: 10 mV

### Alarm Status

Rate Alarm: 200 kcpm                      Rate Alert: 100 kcpm                      Scal Alarm: 915 kcts

### Detector Status

CC: 100e-2                      DT: 0 usec                      Oper. V: 850 V

### Comments

Detector settings & voltage are identical for switch position 1 & 2.

Response to Thorium welding rod check source #114 on contact = 13.7 kcpm ±20% (16.4 kcpm - 11.0 kcpm)

### Statement of Certification

MJW Technical Services, Inc certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology or to the calibrator facilities of other International Standards organization members, or have been derived from accepted values of natural physical constants or have been derived by the ratio type of calibration techniques. The calibration system conforms to the requirements ISO/IEC 17025 and ANSI N323. The Instrument listed above was inspected prior to shipment and it met all the manufacturer's published operating specifications. (MJW Technical Services is not responsible for damage incurred during shipment or use of this instrument).

Calibration Technician: (AL) <i>Alexander</i>	QA Technician: <i>[Signature]</i>	Date: 9-1-15
Calibration Date: 09/01/2015	Calibration Due: 09/01/2016	




# Certificate of Calibration

2015-6241-1

243 Root Street Suite 100  
Olean, NY 14760

Customer		Instrument	
Customer Name: MJW Corporation		Manufacturer: Ludlum Measurements	
Address: 15 Hazelwood Drive Suite 112 Amherst, NY 14228		Model: 2241-2	SN: 206098
Customer PO#		Detector Manufacturer: Ludlum Measurements	
Work Order: 2015-6241		Det. Model: 44-10	SN: PR112642 1-Red
Calibration Method: Electronic(*)		Procedure: SOP-06	
Instrument Received: <input checked="" type="checkbox"/> Within Tolerance <input type="checkbox"/> Out of Tolerance <input type="checkbox"/> Repairs Required <input type="checkbox"/> Other (See Comments)			
<input type="checkbox"/> Geotropism <input type="checkbox"/> Meter Zero <input checked="" type="checkbox"/> Mech. Ck <input type="checkbox"/> HV Readout <input checked="" type="checkbox"/> Battery Check <input checked="" type="checkbox"/> Reset			
<input checked="" type="checkbox"/> Audio <input type="checkbox"/> Window Status <input checked="" type="checkbox"/> FS Response <input checked="" type="checkbox"/> Linearity <input checked="" type="checkbox"/> Alarm Set			
Temperature: 71 F		Humidity: 52%	
		Pressure: 28.5 in Hg	
		Altitude: 1450 ft	

## Source Voltage Plateaus



### Statement of Certification

MJW Technical Services, Inc certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology or to the calibrator facilities of other International Standards organization members, or have been derived from accepted values of natural physical constants or have been derived by the ration type of calibration techniques. The calibration system conforms to the requirements ISO/IEC 17025 and ANSI N323. The Instrument listed above was inspected prior to shipment and it met all the manufacturer's published operating specifications. (MJW Technical Services is not responsible for damage incurred during shipment or use of this instrument).

Calibration Technician: (AL) <i>Adam W. [Signature]</i>	QA Technician: <i>[Signature]</i>	Date: 9-1-15
Calibration Date: 09/01/2015	Calibration Due: 09/01/2016	



# Certificate of Calibration

2015-6243-1

243 Root Street Suite 100  
 Olean, NY 14760

Customer		Instrument	
Customer Name: MJW Corporation		Manufacturer: Ludlum Measurements	
Address: 15 Hazelwood Drive Suite 112 Amherst, NY 14228		Model: 2241-2	SN: 262737
Customer PO#		Detector Manufacturer: Ludlum Measurements	
Work Order: 2015-6243		Det. Model: 44-10	SN: PR111127 1-Red
Calibration Method: Electronic(*)		Procedure: SOP-06	
Instrument Received: <input checked="" type="checkbox"/> Within Tolerance <input type="checkbox"/> Out of Tolerance <input type="checkbox"/> Repairs Required <input type="checkbox"/> Other (See Comments)			
<input type="checkbox"/> Geotropism <input type="checkbox"/> Meter Zero <input checked="" type="checkbox"/> Mech. Ck <input type="checkbox"/> HV Readout <input checked="" type="checkbox"/> Battery Check <input checked="" type="checkbox"/> Reset			
<input checked="" type="checkbox"/> Audio <input type="checkbox"/> Window Status <input checked="" type="checkbox"/> FS Response <input checked="" type="checkbox"/> Linearity <input checked="" type="checkbox"/> Alarm Set			
Temperature: 71.6 F		Humidity: 50%	Pressure: 28.5 in Hg Altitude: 1450 ft

Instrument Calibration									
Multiplier/Range	Calibration Point	Instrument Response			Multiplier/Range	Calibration Point	Instrument Response		
		As Found	As Left	Tolerance			As Found	As Left	Tolerance
Ratemeter Mode*	400 cpm	400 cpm	400 cpm	360-440	Ratemeter Mode*	4 kcpm	3.99 kcpm	3.99 kcpm	3.6-4.4
Ratemeter Mode*	40 kcpm	39.9 kcpm	39.9 kcpm	36-44	Ratemeter Mode*	400 kcpm	399 kcpm	399 kcpm	360-440
Scaler Mode*	400 cpm	400 cpm	400 cpm	360-440	Scaler Mode*	4 kcpm	3.991 kcpm	3.991 kcpm	3.6-4.4
Scaler Mode*	40 kcpm	39.942 kcpm	39.942 kcpm	36-44	Scaler Mode*	400 kcpm	399.824 kcpm	399.824 kcpm	360-440

**Electronic Instruments**

Pulser: 500-2 sn# 220106

Sources								
Isotope	Serial#	Type	Activity	Response	Efficiency	CF	Distance	
Am241	UB992	Y	2344501.240 dpm	134517	5.7 %	17.54	0	
Cs137	UB994	Y	2380863.342 dpm	214047	9.0 %	11.11	0 inch	

**Instrument Status**

Disc/Sens: 10 mV

**Alarm Status**

Rate Alarm: 300 kcpm Rate Alert: 200 kcpm Scal Alarm: 300 kcts

**Detector Status**

CC: 100e-2 DT: 0 usec Oper. V: 1000 V Num: 1

**Comments**

Response to Thorium welding rod check source #116 on contact = 16.8 kcpm ±20% (13.44 kcpm - 20.16 kcpm)

**Statement of Certification**

MJW Technical Services, Inc certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology or to the calibrator facilities of other International Standards organization members, or have been derived from accepted values of natural physical constants or have been derived by the ration type of calibration techniques. The calibration system conforms to the requirements ISO/IEC 17025 and ANSI N323. The Instrument listed above was inspected prior to shipment and it met all the manufacturer's published operating specifications. (MJW Technical Services is not responsible for damage incurred during shipment or use of this instrument).

Calibration Technician: (AL) <i>Adam</i>	QA Technician: <i>Mark</i>	Date: 9-3-15
Calibration Date: 09/02/2015	Calibration Due: 09/02/2016	



# Certificate of Calibration

2015-6243-2

243 Root Street Suite 100  
Olean, NY 14760

Customer		Instrument	
Customer Name: MJW Corporation		Manufacturer: Ludlum Measurements	
Address: 15 Hazelwood Drive Suite 112 Amherst, NY 14228		Model: 2241-2	SN: 262737
		Detector Manufacturer: Ludlum Measurements	
Customer PO#	Work Order: 2015-6243	Det. Model: 43-90	SN: PR277929 2-Yellow
Calibration Method: Electronic(*)		Procedure: SOP-06	

Instrument Received:  Within Tolerance  Out of Tolerance  Repairs Required  Other (See Comments)

Geotropism  Meter Zero  Mech. Ck  HV Readout  Battery Check  Reset

Audio  Window Status  FS Response  Linearity  Alarm Set

Temperature: 71.6 F Humidity: 50% Pressure: 28.5 in Hg Altitude: 1450 ft

### Instrument Calibration

Multiplier/Range	Calibration Point	Instrument Response			Multiplier/Range	Calibration Point	Instrument Response		
		As Found	As Left	Tolerance			As Found	As Left	Tolerance
Ratemeter Mode*	400 cpm	400 cpm	400 cpm	360-440	Ratemeter Mode*	4 kcpm	3.99 kcpm	3.99 kcpm	3.6-4.4
Ratemeter Mode*	40 kcpm	39.9 kcpm	39.9 kcpm	36-44	Ratemeter Mode*	400 kcpm	399 kcpm	399 kcpm	360-440
Scaler Mode*	400 cpm	400 cpm	400 cpm	360-440	Scaler Mode*	4 kcpm	3.991 kcpm	3.991 kcpm	3.6-4.4
Scaler Mode*	40 kcpm	39.942 kcpm	39.942 kcpm	36-44	Scaler Mode*	400 kcpm	399.824 kcpm	399.824 kcpm	360-440

### Electronic Instruments

Pulser: 500-2 sn# 220106

### Sources

Isotope	Serial#	Type	Activity	Response	Efficiency	CF	Distance
Pu239	C7-636	a 4pi	111567.863 dpm	21284	19.1 %	5.24	0 inch

### Instrument Status

Disc/Sens: 10 mV

### Alarm Status

Rate Alarm: 300 kcpm Rate Alert: 200 kcpm Scal Alarm: 300 kcts

### Detector Status

CC: 100e-2 DT: 0 usec Oper. V: 700 V Num: 2

### Comments

Response to Thorium welding rod check source #116 on contact = 80 cpm ±20% (64 cpm - 96 cpm)

### Statement of Certification

MJW Technical Services, Inc certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology or to the calibrator facilities of other International Standards organization members, or have been derived from accepted values of natural physical constants or have been derived by the ratio type of calibration techniques. The calibration system conforms to the requirements ISO/IEC 17025 and ANSI N323. The Instrument listed above was inspected prior to shipment and it met all the manufacturer's published operating specifications. (MJW Technical Services is not responsible for damage incurred during shipment or use of this instrument).

Calibration Technician: (AL)	QA Technician:	Date: 9-3-15
Calibration Date: 09/02/2015	Calibration Due: 09/02/2016	



# Certificate of Calibration

2015-6326-4

243 Root Street Suite 100  
Olean, NY 14760

Customer		Instrument	
Customer Name: MJW Corporation		Manufacturer: Ludlum Measurements	
Address: 15 Hazelwood Drive Suite 112 Amherst, NY 14228		Model: 2241 SN: 196664	
Customer PO# Work Order: 2015-6326		Detector Manufacturer: Ludlum Measurements	
Calibration Method: Electronic(*)		Det. Model: 44-10 SN: PR256142	
Instrument Received: <input checked="" type="checkbox"/> Within Tolerance <input type="checkbox"/> Out of Tolerance <input type="checkbox"/> Repairs Required <input type="checkbox"/> Other (See Comments)		Procedure: SOP-06	
<input type="checkbox"/> Geotropism <input checked="" type="checkbox"/> Meter Zero <input checked="" type="checkbox"/> Mech. Ck <input type="checkbox"/> HV Readout <input checked="" type="checkbox"/> Battery Check <input checked="" type="checkbox"/> Reset			
<input checked="" type="checkbox"/> Audio <input type="checkbox"/> Window Status <input checked="" type="checkbox"/> FS Response <input checked="" type="checkbox"/> Linearity <input checked="" type="checkbox"/> Alarm Set			
Temperature: 70.2 F Humidity: 45%		Pressure: 28.3 in Hg Altitude: 1450 ft	

Instrument Calibration									
Multiplier/Range	Calibration Point	Instrument Response			Multiplier/Range	Calibration Point	Instrument Response		
		As Found	As Left	Tolerance			As Found	As Left	Tolerance
Ratemeter Mode*	40 cpm	40 cpm	40 cpm	36-44	Ratemeter Mode*	400 cpm	400 cpm	400 cpm	360-440
Ratemeter Mode*	4 kcpm	3.99 kcpm	3.99 kcpm	3.6-4.4	Ratemeter Mode*	40 kcpm	39.9 kcpm	39.9 kcpm	36-44
Ratemeter Mode*	400 kcpm	399 kcpm	399 kcpm	360-440	Scaler Mode*	40 cpm	40 cpm	40 cpm	36-44
Scaler Mode*	400 cpm	400 cpm	400 cpm	360-440	Scaler Mode*	4 kcpm	3.992 kcpm	3.992 kcpm	3.6-4.4
Scaler Mode*	40 kcpm	39.907 kcpm	39.907 kcpm	36-44	Scaler Mode*	400 kcpm	398.987 kcpm	398.987 kcpm	360-440

**Electronic Instruments**  
Pulsar: 500-2 sn# 220106      Scaler: 2200 sn# 287163

Sources								
Isotope	Serial#	Type	Activity	Response	Efficiency	CF	Distance	
Am241	UB992	Y	2344079.688 dpm	131062	5.6 %	17.86	0 inch	
Cs137	UB994	Y	2374408.107 dpm	235181	9.9 %	10.10	0 inch	

**Instrument Status**  
Voltage: 1400 V      Disc/Sens: 10 mV

**Alarm Status**  
Rate Alarm: 300 kcpm      Rate Alert: 200 kcpm      Scal Alarm: 300 kcts

**Detector Status**  
CC: 100E-2      DT: 0 usec      Oper. V: 1400 V

**Comments**  
Set for Gross Counting only with 96" cable.

Model 44-10 energy resolution = 11%, acceptable detector energy resolution is <13%

Statement of Certification		
MJW Technical Services, Inc certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology or to the calibrator facilities of other International Standards organization members, or have been derived from accepted values of natural physical constants or have been derived by the ration type of calibration techniques. The calibration system conforms to the requirements ISO/IEC 17025 and ANSI N323. The Instrument listed above was inspected prior to shipment and it met all the manufacturer's published operating specifications. (MJW Technical Services is not responsible for damage incurred during shipment or use of this instrument).		
Calibration Technician: (AL)	QA Technician:	Date: 10/20/15
Calibration Date: 10/15/2015	Calibration Due: 10/15/2016	






# Certificate of Calibration

2015-6326-4

243 Root Street Suite 100  
Olean, NY 14760

Customer		Instrument	
Customer Name: MJW Corporation		Manufacturer: Ludlum Measurements	
Address: 15 Hazelwood Drive Suite 112 Amherst, NY 14228		Model: 2241 SN: 196664	
Customer PO#	Work Order: 2015-6326	Detector Manufacturer: Ludlum Measurements	
Calibration Method: Electronic(*)		Det. Model: 44-10 SN: PR256142	
Instrument Received: <input checked="" type="checkbox"/> Within Tolerance <input type="checkbox"/> Out of Tolerance <input type="checkbox"/> Repairs Required <input type="checkbox"/> Other (See Comments)			
<input type="checkbox"/> Geotropism <input checked="" type="checkbox"/> Meter Zero <input checked="" type="checkbox"/> Mech. Ck <input type="checkbox"/> HV Readout <input checked="" type="checkbox"/> Battery Check <input checked="" type="checkbox"/> Reset			
<input checked="" type="checkbox"/> Audio <input type="checkbox"/> Window Status <input checked="" type="checkbox"/> FS Response <input checked="" type="checkbox"/> Linearity <input checked="" type="checkbox"/> Alarm Set			
Temperature: 70.2 F	Humidity: 45%	Pressure: 28.3 in Hg	Altitude: 1450 ft

## Source Voltage Plateaus



### Statement of Certification

MJW Technical Services, Inc certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology or to the calibrator facilities of other International Standards organization members, or have been derived from accepted values of natural physical constants or have been derived by the ration type of calibration techniques. The calibration system conforms to the requirements ISO/IEC 17025 and ANSI N323. The Instrument listed above was inspected prior to shipment and it met all the manufacturer's published operating specifications. (MJW Technical Services is not responsible for damage incurred during shipment or use of this instrument).

Calibration Technician:  
(AL)

QA Technician:

Date:

10/30/15

Calibration Date: 10/15/2015

Calibration Due: 10/15/2016



# Certificate of Calibration

2015-6361-1

243 Root Street Suite 100  
Olean, NY 14760

Customer		Instrument	
Customer Name: MJW Corporation		Manufacturer: Ludlum Measurements	
Address: 15 Hazelwood Drive Suite 112 Amherst, NY 14228		Model: 2241-2 SN: 262641	
Customer PO# Work Order: 2015-6361		Detector Manufacturer: Ludlum Measurements	
Calibration Method: Electronic(*)		Det. Model: 43-90 SN: PR277930 2-Yellow	
Instrument Received: <input checked="" type="checkbox"/> Within Tolerance <input type="checkbox"/> Out of Tolerance <input type="checkbox"/> Repairs Required <input type="checkbox"/> Other (See Comments)		Procedure: SOP-06	
<input type="checkbox"/> Geotropism <input type="checkbox"/> Meter Zero <input checked="" type="checkbox"/> Mech. Ck <input type="checkbox"/> HV Readout <input checked="" type="checkbox"/> Battery Check <input checked="" type="checkbox"/> Reset			
<input checked="" type="checkbox"/> Audio <input type="checkbox"/> Window Status <input checked="" type="checkbox"/> FS Response <input type="checkbox"/> Linearity <input checked="" type="checkbox"/> Alarm Set			
Temperature: 69.1 F Humidity: 41%		Pressure: 28.6 in Hg Altitude: 1450 ft	

Instrument Calibration									
Multiplier/Range	Calibration Point	Instrument Response			Multiplier/Range	Calibration Point	Instrument Response		
		As Found	As Left	Tolerance			As Found	As Left	Tolerance
Ratemeter Mode*	40 cpm	39.3 cpm	39.3 cpm	36-44	Ratemeter Mode*	400 cpm	400 cpm	400 cpm	360-440
Ratemeter Mode*	4 kcpm	3.96 kcpm	3.96 kcpm	3.6-4.4	Ratemeter Mode*	40 kcpm	39.6 kcpm	39.6 kcpm	36-44
Ratemeter Mode*	400 kcpm	396 kcpm	396 kcpm	360-440	Scaler Mode*	40 cpm	40 cpm	40 cpm	36-44
Scaler Mode*	400 cpm	396 cpm	396 cpm	360-440	Scaler Mode*	4 kcpm	3.968 kcpm	3.968 kcpm	3.6-4.4
Scaler Mode*	40 kcpm	39.677 kcpm	39.677 kcpm	36-44	Scaler Mode*	400 kcpm	396.678 kcpm	396.678 kcpm	360-440

**Electronic Instruments**

Pulsar: 500-2 sn# 220099      Scaler: 2200 sn# 304428

Sources							
Isotope	Serial#	Type	Activity	Response	Efficiency	CF	Distance
Am241	C7-660	a 4pi	110229.264 dpm	25300	23.0 %	4.35	0 inch
Pu239	C7-635	a 4pi	111427.358 dpm	26250	23.6 %	4.24	0 inch
Th230	C7-643	a 4pi	99310.909 dpm	22370	22.5 %	4.44	0 inch

**Instrument Status**

Disc/Sens: 10 mV

**Alarm Status**

Rate Alarm: 800 kcpm      Rate Alert: 500 kcpm      Scal Alarm: 990 kcpm

**Detector Status**

CC: 100e-2      DT: 0 usec      Oper. V: 900 V      Num: 2

Statement of Certification			
<p>MJW Technical Services, Inc certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology or to the calibrator facilities of other International Standards organization members, or have been derived from accepted values of natural physical constants or have been derived by the ration type of calibration techniques. The calibration system conforms to the requirements ISO/IEC 17025 and ANSI N323. The Instrument listed above was inspected prior to shipment and it met all the manufacturer's published operating specifications. (MJW Technical Services is not responsible for damage incurred during shipment or use of this instrument).</p>			
Calibration Technician: (GH)		QA Technician:	Date: 11/3/15
Calibration Date:	11/03/2015	Calibration Due:	11/03/2016



# Certificate of Calibration

2015-6361-2

243 Root Street Suite 100  
 Olean, NY 14760

Customer		Instrument	
Customer Name: MJW Corporation		Manufacturer: Ludlum Measurements	
Address: 15 Hazelwood Drive Suite 112 Amherst, NY 14228		Model: 2241-2	SN: 262641
Customer PO#		Detector Manufacturer: Ludlum Measurements	
Work Order: 2015-6361		Det. Model: 44-10	SN: PR288429 1-Red
Calibration Method: Electronic(*)		Procedure: SOP-06	
Instrument Received: <input checked="" type="checkbox"/> Within Tolerance <input type="checkbox"/> Out of Tolerance <input type="checkbox"/> Repairs Required <input type="checkbox"/> Other (See Comments)			
<input type="checkbox"/> Geotropism <input type="checkbox"/> Meter Zero <input checked="" type="checkbox"/> Mech. Ck <input type="checkbox"/> HV Readout <input checked="" type="checkbox"/> Battery Check <input checked="" type="checkbox"/> Reset			
<input checked="" type="checkbox"/> Audio <input type="checkbox"/> Window Status <input checked="" type="checkbox"/> FS Response <input type="checkbox"/> Linearity <input checked="" type="checkbox"/> Alarm Set			
Temperature: 69.1 F		Humidity: 41%	Pressure: 28.6 in Hg Altitude: 1450 ft

Instrument Calibration									
Multiplier/Range	Calibration Point	Instrument Response			Multiplier/Range	Calibration Point	Instrument Response		
		As Found	As Left	Tolerance			As Found	As Left	Tolerance
Ratemeter Mode*	40 cpm	39.3 cpm	39.3 cpm	36-44	Ratemeter Mode*	400 cpm	400 cpm	400 cpm	360-440
Ratemeter Mode*	4 kcpm	3.96 kcpm	3.96 kcpm	3.6-4.4	Ratemeter Mode*	40 kcpm	39.6 kcpm	39.6 kcpm	36-44
Ratemeter Mode*	400 kcpm	396 kcpm	396 kcpm	360-440	Scaler Mode*	40 cpm	40 cpm	40 cpm	36-44
Scaler Mode*	400 cpm	396 cpm	396 cpm	360-440	Scaler Mode*	4 kcpm	3.968 kcpm	3.968 kcpm	3.6-4.4
Scaler Mode*	40 kcpm	39.677 kcpm	39.677 kcpm	36-44	Scaler Mode*	400 kcpm	396.678 kcpm	396.678 kcpm	360-440

**Electronic Instruments**  
 Pulsar: 500-2 sn# 220099 Scaler: 2200 sn# 304428

Sources							
Isotope	Serial#	Type	Activity	Response	Efficiency	CF	Distance
Am241	UB992	Y	2343863.800 dpm	176266	7.5 %	13.33	0 inch
Cs137	UB994	Y	2371561.372 dpm	268517	11.3 %	8.85	0 inch

**Instrument Status**  
 Disc/Sens: 10 mV

**Alarm Status**  
 Rate Alarm: 500 kcpm Rate Alert: 300 kcpm Scal Alarm: 900 kcpm

**Detector Status**  
 CC: 100e-2 DT: 0 usec Oper. V: 1000 V Num: 1

**Comments**  
 Model 44-10 energy resolution = 8.9%, acceptable detector energy resolution is <13%

Statement of Certification		
MJW Technical Services, Inc certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology or to the calibrator facilities of other International Standards organization members, or have been derived from accepted values of natural physical constants or have been derived by the ratio type of calibration techniques. The calibration system conforms to the requirements ISO/IEC 17025 and ANSI N323. The Instrument listed above was inspected prior to shipment and it met all the manufacturer's published operating specifications. (MJW Technical Services is not responsible for damage incurred during shipment or use of this instrument).		
Calibration Technician: (GH)	QA Technician:	Date: 11/3/15
Calibration Date: 11/03/2015	Calibration Due: 11/03/2016	




# Certificate of Calibration

2015-6361-2

243 Root Street Suite 100  
Olean, NY 14760

Customer		Instrument	
Customer Name: MJW Corporation		Manufacturer: Ludlum Measurements	
Address: 15 Hazelwood Drive Suite 112 Amherst, NY 14228		Model: 2241-2	SN: 262641
Customer PO#		Detector Manufacturer: Ludlum Measurements	
Work Order: 2015-6361		Det. Model: 44-10	SN: PR288429 1-Red
Calibration Method: Electronic(*)		Procedure: SOP-06	
Instrument Received: <input checked="" type="checkbox"/> Within Tolerance <input type="checkbox"/> Out of Tolerance <input type="checkbox"/> Repairs Required <input type="checkbox"/> Other (See Comments)			
<input type="checkbox"/> Geotropism <input type="checkbox"/> Meter Zero <input checked="" type="checkbox"/> Mech. Ck <input type="checkbox"/> HV Readout <input checked="" type="checkbox"/> Battery Check <input checked="" type="checkbox"/> Reset			
<input checked="" type="checkbox"/> Audio <input type="checkbox"/> Window Status <input checked="" type="checkbox"/> FS Response <input type="checkbox"/> Linearity <input checked="" type="checkbox"/> Alarm Set			
Temperature: 69.1 F		Humidity: 41%	
		Pressure: 28.6 in Hg	
		Altitude: 1450 ft	

## Source Voltage Plateaus



### Statement of Certification

MJW Technical Services, Inc certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology or to the calibrator facilities of other International Standards organization members, or have been derived from accepted values of natural physical constants or have been derived by the ration type of calibration techniques. The calibration system conforms to the requirements ISO/IEC 17025 and ANSI N323. The Instrument listed above was inspected prior to shipment and it met all the manufacturer's published operating specifications. (MJW Technical Services is not responsible for damage incurred during shipment or use of this instrument).

Calibration Technician:  
(GH)

QA Technician:

Date: 11/3/15

Calibration Date: 11/03/2015

Calibration Due: 11/03/2016